

Figure 1. Physical Photo of A1UHVACP5KV10MARM

FEATURES

High precision

High efficiency

High output voltage stability

Linear modulation of output voltage

Overcurrent protection

Arc and Short circuit protection

Digital display for output voltage and current

OEM Customization Available

APPLICATIONS

A1UHVACP5KV10MARM, is designed for achieving AC-DC conversion from AC voltage to high DC voltage. This series high voltage power supply is rack-mounted, and is mainly used in ion beam injection, semiconductor technology, electron beam welding, capacitor charging, high power radio frequency transmitter, accelerator, CW laser, high voltage insulation test, lithography technology, X-ray system, electrostatic applications, scientific experiments, industrial applications and other fields.

DESCRIPTION

The power supply comes with high stability, high reliability, high accuracy of output voltage regulation, and continuously adjustable output voltage; low time drift and temperature drift; it adopts modular design inside the power supply, with

protection against overvoltage, overcurrent, and arcing; the control interface uses industrial interface, with industrial design from circuit schematic to the overall structure; high integration and high detection accuracy; it has functions such as start-up soft-start and control; convenient for control, easy operation; the internal transformation of the power supply can be made based on the actual needs of the customer

SAFETY PRECAUTIONS

High voltage power supply must be connected to ground reliably.

Do not touch the high voltage wire, unless the high voltage power supply is powered off, and the load and internal capacitors are fully discharged.

When the high voltage power supply is powered off, wait for another 5 minutes for fully discharging all the capacitors inside the power supply.

Do not operate the power supply in humid environment, and do not connect the operator to ground.

The internal protection circuit is provided in the high voltage power supply, but the high voltage short circuit shall be avoided.

Make sure the circuit is insulated perfectly, especially between the high voltage output and the surroundings so as to avoid electronic shock.



SPECIFICATIONS

Table 1. Characteristics. $T_A = 25$ °C, unless otherwise noted

Para	meter	Symbol	Condition	Min.	Тур.	Max.	Unit/Note
AC Inpu	ıt Voltage	$V_{ m VPS}$		100		230	V _{AC}
Quiascant I	Quiescent Input Current		$I_{OUT} = 0 \text{mA}$ $V_{VPS} = 110 \text{ V}_{AC}$	100	150	200	mA
Quiescent	input Current	I_{INQQ}	$I_{OUT} = 0mA$ $V_{VPS} = 220V_{AC}$	65	75	85	mA
Eull Lood I	and Comment		$I_{OUT} = 10 \text{mA}$ $V_{VPS} = 110 V_{AC}$	0.5	0.6	0.7	A
Full Load I	nput Current	${ m I}_{ m INFLD}$	$I_{OUT} = 10 \text{mA}$ $V_{VPS} = 220 V_{AC}$	0.2	0.3	0.4	A
Outpu	t Voltage	V_{OUT}	$I_{OUT} = 0 \sim 10 \text{mA}$	0		5000	V
Maximum C	Output Current	I_{OUTMAX}	Full load			10	mA
Ri	pple		Bandwidth=1MHz		<0.1		%V _{P-P}
L	oad				500		ΚΩ
Dotantiamat	er Adjustment			10k potentiometer or			
Fotentiomet	ei Aujustinent			remote	control () ∼ +5V	
Output Modu	lation Linearity				< 0.1		%
Voltage	Load		$I_{OUT} = 0 \sim 10 \text{mA}$		< 0.1		%
Regulation	Input		$V_{VPS} = 110V \sim 130V$		<±0.1		%
Current	Load		$I_{OUT} = 0 \sim 10 \text{mA}$		< 0.1		%
Regulation	Input		$V_{VPS} = 110V \sim 130V$		< 0.1		%
Full Load	Efficiency	η			≥75		%
Temperatur	e Coefficient	TCV_{O}	0 ~ 50°C		< 0.1		%/°C
T. D.G	Short Time Drift		After 30 minute's		< 0.1		%/ min
Time Drift	Long Time Drift		warming up		< 0.5		%/h
Output Voltage Te	emperature Stability		0 ~ 50°C		<±0.1		%
Operating Te	mperature Range	Topr		0		50	°C
Storage Temp	Storage Temperature Range			-20		85	°C
Humidity				20		elative hu	-
External Dimensions				482.5×350×44 mm		mm	
				19×13.78×1.73 inc		inch	
***	-1-1-4				3.6		kg
We	eight				7.93		lbs

		126.98	Oz
			_

PANEL INSTRUCTIONS

Front Panel




Figure 2. Front Panel

Table 2. Definitions of front panel keys

No.	Signal	Signal Parameters	
1	Power	When the power is turned on, the indicator is on; when the power is turned off, the indicator is off.	
2	HV Switch	High voltage ON, the indicator is on.	
3	R/L	When the switch is in Remote, it is controlled remotely; when the switch is in Local, is controlled locally.	
4	10-turn potentiometer for adjusting output voltage. Rotate it clockwise to in output voltage, and the potentiometer resistance = the corresponding scale > example, when the scale is 10, and the frame above the scale shows 1 (1kΩ resistance = $10 \times 10\Omega + 1k\Omega = 1.1k\Omega$, and the like. HV output: 1.5m long connection wire outputs 5kV.		
5	Constant Limit Indicator	When the voltage exceeds the preset voltage, the indicator is on.	
6	Output Voltage Display	Digital display for the output voltage. The actual output voltage = the reading \times 10V.	
7	Constant Limit Indicator	When the current exceeds the preset current, the indicator is on.	
8	Output Current Display Digital display for the output current. The actual output current = the reading × n		
9	Current Adjustment	10-turn potentiometer for adjusting output voltage. Rotate it clockwise to increase the output voltage, and the potentiometer resistance = the corresponding scale \times 10 Ω . For example, when the scale is 10, and the frame above the scale shows 1 (1k Ω), then the	

	resistance = $10 \times 10\Omega + 1 k\Omega = 1.1 k\Omega$, and the like.
	HV output: 1.5m long connection wire outputs 10mA.

Back Panel




Figure 3. Front Panel

Table 3. Definitions of back panel function keys

No.	Signal	Signal Parameters	
10	HV Output	High voltage output port with removable 1m long connection wire outputting 5kV and 10mA.	
11	Output Ground	High voltage power supply output ground terminal.	
12	Remote	8-pin connector. Refer to the following Table 4 and Figure 4 for remote control function.	
13	Ac Input Voltage Switch	110V or 220V Input voltage selector switch	
14	AC Main Power Input	AC input 110V\220V 50/60Hz connector.	




Figure 4. 8-pin Connector

Table 4. 8-pin connector.

Pin	Signal Signal Parameters		Lead the color
P1	Remote Voltage Program In	0 to 5V=0 to 100% rated output.	White
P2	Voltage Monitor	0 to 5V=0 to 100% rated output.	Green
Р3	Current Monitor	0 to 5V=0 to 100% rated output.	Blue

P4	Remote Current Program In	0 to 5V=0 to 100% rated output.	Yellow
P5	Shutdown	A logic low 0.8V or a 0V on the SDN pin will turn the device off.	Brown
		When SDN is in logic high 1.2V <sdn<5v left="" or="" td="" unconnected,<=""><td></td></sdn<5v>	
		the product is working well.	
P6	Voltage Reference	+5VDC Reference	Red
P7	GND	Signal ground	Black
P8	NC	No connection	

TESTING DATA

High voltage power supply testing data (Test condition: the load is $500\text{K}\Omega$).




Figure 5. V_{CTRL} vs. V_{OUT}

NAMING INSTRUCTIONS




Figure 6. Naming Rules of A1UHVACP5KV10MARM



DIMENSIONS

I. Dimension of the leads.

Leads	Diameter (mm)	Length (m)
High voltage output wire	4.5	1.5
Power cord	6.5	1.8

II. Dimension of A1UHVACP5KV10MARM.




Figure 7. Dimensions for A1UHVACP5KV10MARM

High Voltage Power Supply



A1UHVACP5KV10MARM

NOTICE

- 1. ATI warrants performance of its products for one year to the specifications applicable at the time of sale, except for those being damaged by excessive abuse. Products found not meeting the specifications within one year from the date of sale can be exchanged free of charge.
- 2. ATI reserves the right to make changes to its products or to discontinue any product or service without notice, and advise customers to obtain the latest version of relevant information to verify, before placing orders, that information being relied on is current and complete.
- 3. All products are sold subject to the terms and conditions of sale supplied at the time of order acknowledgment, including those pertaining to warranty, patent infringement, and limitation of liability. Testing and other quality control techniques are utilized to the extent ATI deems necessary to support this warranty. Specific testing of all parameters of each device is not necessarily performed, except those mandated by government requirements.
- 4. Customers are responsible for their applications using ATI components. In order to minimize risks associated with the customers' applications, adequate design and operating safeguards must be provided by the customers to minimize inherent or procedural hazards. ATI assumes no liability for applications assistance or customer product design.
- 5. ATI does not warrant or represent that any license, either express or implied, is granted under any patent right, copyright, mask work right, or other intellectual property right of ATI covering or relating to any combination, machine, or process in which such products or services might be or are used. ATI's publication of information regarding any third party's products or services does not constitute ATI's approval, warranty or endorsement thereof.
- 6. IP (Intellectual Property) Ownership: ATI retains the ownership of full rights for special technologies and/or techniques embedded in its products, the designs for mechanics, optics, plus all modifications, improvements, and inventions made by ATI for its products and/or projects.