NEVO+1200M

MEDICAL DATA SHEET

AC/DC Modular Configurable PSU



1200W

Powerful

6" x 6" x 1.61"

Small

1.2kg

Light

The Ultimate 1200 Watt Configurable Solution

The NEVO+1200M is the smallest in its class and the ultimate power solution for medical applications where size, weight, low standby power and primary side inhibit are vital factors and delivers up to 1200 Watts from a 1.2kg 6" x 6" x 1.61" package. Each configured unit consists of an input module with up to eight output modules, where any combination of outputs can be fitted to create a power solution with up to sixteen isolated outputs. Standard features include intelligent fan control, wide output voltage adjust capability and primary side shutdown with standby power consumption of less than 3 Watts. A low noise fan option with virtually silent operation is also available, which allows you to use this innovative power supply in even the quietest of environments. The series carries full IEC/UL60601-1 3rd edition & IEC/UL60601-1-2 4th edition safety approvals, complies with EN61000 Immunity, EN55022-B EMC Standards and features market leading specifications and design in application support.

MAIN FFATURES

 Up to 1200 Watts of output power 	 IEC/UL60601-1 Ed. 3 & -1-2 Ed. 4 (EMC) 	 Accurate current sharing
 Primary side remote on/off function 	 Industry leading power density (21W/in³) 	 Parallel and series connection of modules
• Standby power ≤ 3 Watts	 Lightest modular design – only 1.2kg – 1000Watts/kg 	• 2 x 5V 1A bias supply
6" x 6" x 1.61" footprint	 Efficiency up to 89% 	 Field configurable
 Low noise fan option 	 Remote current / voltage programming 	 RoHS compliant
		3 Year warranty

APPLICATIONS

 Test & Measurement equipment 	 Laboratory & Analysis equipment 	 LED lighting
 Robotics 	Display	 Retrofit of legacy PSUs
• Oil & Gas	 Avionics 	• Lasers
 Telecommunications 		

CUSTOMER BENEFITS

Fast time to market	 Proven technology 	 Technology consolidation
 24 hrs samples from distribution 	 Eliminates custom design costs 	 Supplier consolidation
 Safety & EMC certified 	 Field replaceable 	
 World class engineering support 	 Low cost of ownership 	

SPECIFICATIONS

INPUT MODULE SPECIFICATIONS						
Parameter	Details	Min	Typical	Max	Units	
AC Input Voltage	Nominal range is 100V _{RMS} to 240V _{RMS}	85		264	V_{RMS}	
AC Input Frequency	Contact factory for 400Hz operation.	47	50/60	63	Hz	
DC Input Voltage	Not covered by safety approvals. Contact Vox Power.	120		370	V_{DC}	
Output Power Rating	De-rate linearly from 1200Watts at 120V _{RMS} to 850Watts at 85V _{RMS}			1200	Watts	
Input Current	1200Watts output at 120V _{RMS} input			12	Amps	
Input Current Limit	Maintains power factor		14		Amps	
Inrush Current	265V _{RMS} , 25°C (cold start)			40	Amps	
Fusing	Live line fused (5x20 Fast acting)			12.5	Amps	
Efficiency	See graphs		86	89	%	
No load Power consumption	All outputs fitted and disabled/enabled		32/46		Watts	
Standby Power	Latched off state, 120Vrms		2.5		Watts	
Power Factor			0.96	0.99		
Holdup	1200Watts output at 120V _{RMS} input	17	20	21	mS	
UVP	Turn on under voltage protection	78		84	V_{RMS}	
Over temperature	Internally monitored.	115		125	°C	
Reliability (1)	Input module	•		1.62	FPMH	
	Fan (2 Fans per unit)			2.7	FPMH	
Warranty	Standard terms and conditions apply			3	Years	
Size 154.5 (L) x 152.4 (W) x 41.0 (H). See diagram for tolerance details					mm	
Weight	720 + 60 per output module				Grams	
Note 1.	30°C base & ambient, 100% load, SR332 Issue 2 Method I, Case 3, Ground, Fixed, Cont	rolled				

GLOBAL SIGNALS SPECIFICATIONS						
Parameter	Details	Min	Typical	Max	Units	
Bias Voltage	Two isolated Bias Outputs available	4.8	5	5.2	Volts	
Bias Current	Hiccup type current limit	0		1	Amps	
AC_OK Voltage	Low output level/High output level	0/3.5	0.2/4.5	1/5.2	Volts	
AC_OK Current		-10		20	mA	
Power Good Voltage	PNP open collector with internal 10kΩ pull down. Low output level/High output level	0/8	0/10	0/15	Volts	
Power Good Current	Open collector output. Current source only. All Slots.			20	mA	
Global Inhibit Voltage	Low input level/High input level.	0/3		1/15	Volts	
Global Inhibit Current	5k input impedance.	0.6		3	mA	
Inhibit Voltage	Low input level/High input level. All slots.	0/2.5		1/15	Volts	
Inhibit Current	10k input impedance. All slots.	0.25		1.5	mA	
Primary Bias voltage	Medically Isolated	4.8	5	5.2	Volts	
Primary Bias current	Hiccup type current limit			0.5	Amps	
Primary Remote On/Off	Negative Edge Triggered, Refer to User Manual		5		Volts	

	OUTPUT MODULE SPECIFICATION SUMMARY											
MODEL	Out Min.	put Volta Nom.	age Max.	Output Current	Rated Power	Peak Power	Load Reg.	Line Reg.	Cross Reg.	Ripple & Noise	FPMH (1)	Feature Set (2)
OD1											0.5	
OP1	1.5V	5V	7.5V	25A	125W	187.5W	±50mV	±5mV	±10mV	50mV _{PP}	0.5	ABCDEFG
OP2	4.5V	12V	15V	15A	150W	225W	±100mV	±12mV	±24mV	120mV_{PP}	0.5	ABCDEFG
OP3	9V	24V	30V	7.5A	150W	225W	±150mV	±24mV	±48mV	240mV _{PP}	0.5	ABCDEFG
OP4	18V	48V	58V	3.75A	150W	217.5W	±300mV	±48mV	±96mV	480mV _{PP}	0.5	ABCDEFG
OP5	3.3V	12V	15V	5A	2x 75W	2x 75W	±50mV	±12mV	±24mV	240mV _{PP}	0.75	AFG
OP8	23.2V	24V	24.7V	3.125A	2x 75W	2x 75W	±100mV	±24mV	±48mV	480mV _{PP}	0.75	AFG
OPA2	4.5V	12V	15V	25A	300W	375W	±100mV	±12mV	±24mV	120mV _{PP}	0.5	ABCDEFGH
OPA3	9V	24V	30V	15A	300W	450W	±150mV	±24mV	±48mV	240mV _{PP}	0.5	ABCDEFGH

Note 1. Output module, 30°C base, 100% load, SR332 issue 2 Method I, Case 3, Ground, Fixed, Controlled

Note 2. A = Remote Sense, B = External Voltage control, C = External constant current control, D = Current output signal, E = Current share, F = Over Voltage protection,
G = Over temperature protection, H = Dual Slot module

SAFETY SPECIFICATIONS					
Parameter	Details	Max	Units		
	Input to Output (2 MOPP). Do not perform test on assembled unit ⁽¹⁾	4000	V_{AC}		
Landa Cara Walter and	Input to Chassis (1 MOPP)	1500	V_{AC}		
Isolation Voltages	Global signals (J2) to Output/Chassis	250	V_{DC}		
	Output to Output/Chassis (Standard modules)	250	V_{DC}		
Earth Leakage Current	Normal condition, 264Vac, 63Hz, 25°C	300	uA		
Touch Leakage Current	Standard modules NC/SFC	20/200	uA		
Patient Leakage Current	Standard modules 264Vac, 63Hz, 25°C NC/SFC ⁽²⁾		uA		
Note 1. Testing an assembled unit to $4000V_{AC}$ may cause damage. Please refer to application note (APN-002) on Vox Power website or contact Vox Power representative.					
Note 2. Not Applicable					

INSTALLATION SPECIFICATIONS							
Parameter Details Parameter Details							
Equipment class	I	Flammability Rating	94V-2				
Overvoltage category	II .	Ingress protection rating	IP10				
Material Group	IIIb (indoor use only)	ROHS compliance	2011/65/EU				
Pollution degree	2	Intended usage environment	Home Healthcare				

ENVIRONMENTAL SPECIFICATIONS						
Dayanaatay			erational	Operational		I Indian
Parameter	Details	Min	Max	Min	Max	- Units
Air Temperature	Operational limits subject to appropriate de-ratings	-40	+85	-20	70	°C
Humidity	Relative, non-condensing	5	95	5	95	%
Altitude		-200	5000	-200	3000	m
Air Pressure		52	106	69	106	kPa
Noise Level	Variable. Measured 1m from fan intake.	-	-	42	61	dBA
Shock	3000 bumps at 10G (16ms) half sine wave					
Vibration	1.5G 10 to 200Hz sine wave, 20G for 15min in 3 axes random vibration					

ELECTROMAGNETIC COMPLIANCE – EMISSIONS Phenomenon Basic EMC Standard Test Details Radiated emissions, electric field EN55011/22, FCC Class A compliant (See note for Class B) Conducted emissions EN55011/22, FCC part 15, CISPR 22/11 Class B compliant Harmonic Distortion IEC61000-3-2 Compliant Flicker & Fluctuation IEC61000-3-3 Compliant


Note: To meet Class B radiated emissions the end user should add ferrites to I/P and O/P cables. Consult Vox Power for details.


ELECTROMAGNETIC COMPLIANCE – IMMUNITY						
Phenomenon	Basic EMC Standard	Test Details				
Electrostatic discharge	IEC61000-4-2	Test level 4: 15kV air, 8kV contact				
Radiated RF EM fields	IEC61000-4-3	Test Level 3: (10V/m, 80MHz-2.7GHz) sine wave AM 80% 1kHz				
Proximity fields from RF wireless communications equipment	IEC61000-4-3	Test levels as per IEC60601-1-2:2014 Table 9				
Electrical Fast Transients/bursts	IEC61000-4-4	Test Level 3: (2kV Power, 1kV I/O) 5kHz(ed3) & 100kHz(ed4)				
Surges	IEC61000-4-5	Test Level 3: 1kV L-N, 2kV L-E				
Conducted disturbances induced by RF fields	IEC61000-4-6	Test Level 3: 10V, 0.15 to 80Mhz sine wave AM 80% 1kHz				
Power Frequency Magnetic Fields	IEC61000-4-8	Test level 4: 30A/m 50Hz				
Voltage Dips	IEC61000-4-11& SEMI-F47-0706 (2)	0% 10ms, 0% 20ms, 80% 1s, 80% 10s, 90% continuous (Criterion A) 70% 0.5s, 40% 0.2s (Criterion A at 240V and Criterion B at 100V)				
Voltage interruptions	IEC61000-4-11	0% 250/300 cycle as per IEC60601-1-2:2014 (Criterion B)				


Notes:
1. Criterion A = No degradation of performance or loss of function.
Criterion B = Temporary degradation of performance or loss of function is allowed, provided the function is self-recoverable.
Criterion C = Temporary loss of function is allowed but requires operator intervention to recover.


2. Tested at nominal range (100V to 240V). Line deratings applied where appropriate.


AGENCY APPROVALS						
Standard	Details	File				
IEC 60601-1:2005 + CORR1 2006 + CORR2: 2007 + A1:2012	Medical electrical equipment Part 1: General requirements for basic safety and essential performance	UL: E316486				
EN60601-1:2006 + A11:2011 + A1:2013 + A12:2014	Medical electrical equipment Part 1: General requirements for basic safety and essential performance					
CAN/CSA-C22.2 No. 60601-1 (2008)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential Performance					
ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)	Medical Electrical Equipment Part 1: General Requirements for Basic Safety and Essential Performance					
CE MARK	LVD 2014/35/EU, EMC 2014/30/EU					
CB certificate and report available on request						




















All specifications are believed to be correct at time of publishing. Vox Power Ltd reserves the right to make changes to any of its products and to change or improve any part of the specification, electrical or mechanical design or manufacturing process without notice. Vox Power Ltd does not assume any liability arising out of the use or application of any of its products and of any information to the maximum extent permitted by law. No license, express or implied, by estoppel or otherwise, to any intellectual property rights is granted by this document or by any products of Vox Power Ltd. VOX POWER LTD DISCLAIMS ALL WARRANTIES AND REPRESENTATIONS OF ANY KIND WHETHER EXPRESS OR IMPLIED, INCLUDING, BUT NOT LIMITED TO, IMPLIED WARRANTIES FOR PURPOSE, MERCHANTABILITY AND NONINFRINGEMENT. Please consult your local distributor or Vox Power directly to ensure that you have the latest revision before using the product and refer to the latest relevant user manual for further information relating to the use of the product. Vox Power Ltd products are not intended for use in connection with life support systems, human implantations, nuclear facilities or systems, aircraft, spacecraft, military or naval missile, ground support or control equipment used for the purpose of guidance navigation or direction of any aircraft, spacecraft or military or naval missile or any other application where product failure could lead to loss of life or catastrophic property damage. The user will hold Vox Power Ltd harmless from any loss, cost or damage resulting from its breach of these provisions.