

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

Product

Name and address of the applicant

Name and address of the manufacturer

Name and address of the factory

Note: When more than one factory, please report on page 2

Ratings and principal characteristics

Trademark (if any)

Type of Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Additional information (if necessary may also be reported on page 2)

A sample of the product was tested and found to be in conformity with

As shown in the Test Report Ref. No. which forms part of this Certificate

VCCM600M switch mode modular power supplies for installation into medical equipment
VOX POWER LTD
Unit 2
Redcow Interchange Estate
Dublin, 22 Ireland

VOX POWER LTD
Unit 2
Redcow Interchange Estate
Dublin, 22 Ireland

EOS POWER INDIA PVT LTD
UNIT 57 SDF-II
SEEPZ ANDHERI (E)
MUMBAI, 400096 India

[Additional Information on page 2](#)
See Page 2



VCCM600M
See Page 2

Additionally evaluated to EN 60601-1: 2006 / A1: 2013 / A12: 2014; National Differences specified in the CB Test Report.

[Additional Information on page 2](#)
IEC 60601-1(ed.3), IEC 60601-1(ed.3);am1

E316486-D1000-1/A0/C0-ULCB issued on 2017-03-16

This CB Test Certificate is issued by the National Certification Body



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-03-22

Signature:

Jan-Erik Storgaard



Model Details:

VCCM600M May be followed by: -abcd

Where:

- a (Slot 1) = 0, A, B, C, D;
- b (Slot 2) = 0, A, B, C, D;
- c (Slot 3) = 0, A, B, C, D;
- d (Slot 4) = 0, A, B, C, D; and

where 0 = No Module Fitted; and
 where A = OPA; B = OPB; C = OPC and D = OPD

May also be followed by four alphanumeric characters (-zzzz) to indicate non-safety information.

Output Modules: OPA; OPB; OPC; OPD

Factories:

Panyu Trio Microtronics Co Ltd
 SHIJI INDUSTRIAL ESTATE
 DONGYONG
 NANSHA
 GUANGZHOU, 511453 GUANGDONG China

Ratings:

VCCM600M input rating:

- 100-240 V ac
- (or 120-370 V dc)
- 7 A, 50-60 Hz

Maximum output power is 600 W.

Individual outputs rated according to modules fitted.

Output modules output ratings (optional):

- OPA = 1.5 - 7.5 V dc, 25 A, 125 W
- OPB = 4.5 - 15 V dc, 15 A, 150 W
- OPC = 9 - 30 V dc, 7.5 A, 150 W
- OPD = 18 - 58 V dc, 3.75 A, 150 W

All output modules current and power ratings are the maximum allowable.

Additional information (if necessary)

Information complémentaire (si nécessaire)



- UL (US), 333 Pfingsten Rd IL 60062, Northbrook, USA
- UL (Demko), Borupvang 5A DK-2750 Ballerup, DENMARK
- UL (JP), Marunouchi Trust Tower Main Building 6F, 1-8-3 Marunouchi, Chiyoda-ku, Tokyo 100-0005, JAPAN
- UL (CA), 7 Underwriters Road, Toronto, M1R 3B4 Ontario, CANADA

For full legal entity names see www.ul.com/ncbnames

Date: 2017-03-22

Signature:

Jan-Erik Storgaard