MALLORY Mallory Sonalert Products,Inc. Sales Outline Drawing			Part #: SBT5LMMPPC	
			Revision:	F
	1	alli Dagamiradi	- IFCC0C04 4 0	
Specifications:	cUL Recognized to IEC60601-1-8			
	Artificial Perfusion Melody			
Sound Level Category:	Silence	Loud	Loud	Loud
Mode of Operation:	Silence - NO SOUND	Low Priority Alarm - 2 beeps repeating every 30 sec	Medium Priority Alarm: 3 beeps repeating every 7.5 sec	High Priority Alarm - 10 beeps repeating every 2.5 sec
Loudness @ 10cm:	Silence	85 to 95 dB(A) Typ.	85 to 95 dB(A) Typ.	85 to 95 dB(A) Typ.
Frequency:	Silence	150 to 4000 Hz	150 to 4000 Hz	150 to 4000 Hz
Connection:	Connect "+" to " + " & " - " to " - " of Power Supply	Connect "+" to "+" & "-" & L to "-" of Power Supply	Connect "+" to "+" & "-" & M to "-" of Power Supply	Connect "+" to "+" & "-" & H to "-" of Power Supply
Voltage Rating:	3.3 to 5 VDC			
Average Current Draw:	<15mA Typ.	<15mA Typ.	<15mA Typ.	<20mA Typ.
Ipk Current*:	250mA			
Housing Material:	ABS 777D Plastic Resin, Black			
Storage Temperature:	-40° to +85° C			
Operating Temperature:	-30° to +65° C			
Weight Typical:	0.6 oz (17g)			
Options:	Please contact factory.			
Dimensions: Inches [m	-	cUL Recognized NUMBER: 7880593	R	OHS Compliant
	Ø1.750 [Ø44.5]	0.334 [8.5]	

0.100 [2.5] 0.200 [5.1] 0.334 [8.5] 0.000 [5.1] 0.000 [5.1]

4411 South High School Road, Indianapolis, Indiana 46241 U.S.A \cdot Phone: (317) 612-1000 Fax: (317) 612-1010

*Ipk = Instantaneous Peak

www.mspindy.com

tolerances:

±0.01 [±0.3]

±0.005 [±0.1]